

PHILIPS

Computed
Tomography

Incisive CT



Intellect at every step

Incisive CT 64 Essentials specifications

Contents

1	Introduction	3	9	Clinical enhancements	9
9.1	Bolus tracking (optional)		9.1	Bolus tracking (optional)	
9.2	Spiral Auto Start (SAS) (optional)		9.2	Spiral Auto Start (SAS) (optional)	
9.3	Dental application (optional)		9.3	Dental application (optional)	
9.4	Clinical applications, standard		9.4	Clinical applications, standard	
9.5	Brain Perfusion (optional)		9.5	Brain Perfusion (optional)	
9.6	Lung Nodule Assessment (optional)		9.6	Lung Nodule Assessment (optional)	
9.7	CT Colonoscopy (optional)		9.7	CT Colonoscopy (optional)	
9.8	Vessel Analysis (optional)		9.8	Vessel Analysis (optional)	
9.9	Cardiology Solutions (optional)		9.9	Cardiology Solutions (optional)	
9.10	Interventional Solutions (optional)		9.10	Interventional Solutions (optional)	
9.11	Dual Energy (optional)		9.11	Dual Energy (optional)	
2	Gantry	4	10	Networking and storage	10
10.1			10.1	Networking	
10.2			10.2	DICOM	
10.3			10.3	DICOM connectivity	
10.4			10.4	DICOM DVD/CD writer	
10.5			10.5	Filming	
3	Patient table	4	11	Site planning	10
11.1			11.1	Power requirements	
11.2			11.2	Console Uninterrupted Power Supply (UPS) (optional)	
11.3			11.3	Environmental requirements	
11.4			11.4	System requirements	
11.5			11.5	Dimensions and weights, per unit (without package)	
4	Accessories	4			
5	Imaging chain	6			
5.1	Generator				
5.2	X-ray tube				
5.3	NanoPanel Elite detector				
6	Image quality	7			
6.1	Spatial resolution				
6.2	Low-contrast resolution				
6.3	Other				
7	Reconstruction	8			
7.1	Reconstruction speed				
7.2	iDose ⁴ (optional)				
7.3	O-MAR				
7.4	Cone Beam Reconstruction Algorithm				
7.5	ClearRay reconstruction				
7.6	Adaptive filtering				
7.7	Adaptive multi-cycle reconstruction (optional)				
7.8	Reconstruction field of view				
7.9	Image matrix				
7.9	Off-line reconstruction				
8	DoseWise	8			
8.1	DoseRight Index				
8.2	Spiral Auto Start (SAS)				
8.3	DICOM structured reporting/IHE REM profile				
8.4	DoseRight automatic current selection				
8.5	DoseRight angular dose modulation				
8.6	3D-DOM				
8.7	Dedicated pediatric protocols				
8.8	Dose display and reports				
8.9	Dose performance data				

1. Introduction

Philips Incisive CT 64 Essentials helps you meet some of your organization’s most pressing challenges. Incisive CT 64 Essentials offers intellect at every step, from acquisition through results, and across all fronts: financial, clinical and operational. Like never before, operator and design efficiencies come together for wise decisions from start to finish.

By offering smart approaches from the start, Incisive CT 64 Essentials:

Elevates your business	Delivers intelligence that adapts to you	Brings predictability to an unpredictable world
<p>Reduce operational costs to meet financial objectives</p> <ul style="list-style-type: none"> • Tube for Life guarantee* • Technology Maximizer • Expansive in-room upgradeability 	<p>Speed workflow and expand clinical breadth</p> <ul style="list-style-type: none"> • OnPlan patient-side gantry controls • IntelliSpace Portal • iDose⁴ and O-MAR • 70 kV scanning 	<p>Reduce downtime and improve efficiency and care</p> <ul style="list-style-type: none"> • Proactive monitoring • Remote services • DoseWise Portal • PerformanceBridge



Feature	Specification
Effective power with iDose ⁴	72 kW, 94 kW**
Slices	64
Coverage	20 mm
Maximum scannable range	1,860 mm
Bore size	720 mm
Reconstruction speed	Up to 28 IPS
Anode effective heat capacity	≥25 MHU _{eff} ; direct cooling

Note: Effective power is calculated by using generator power (55 kW, 72kW**) and using iDose⁴ at the same time. This gives Incisive CT 64 Essentials effectively more power.

* Life of product (or CT system) is defined by Philips as 10 years. Tube for Life guarantee availability varies by country. Please contact your local Philips sales representative for details.

** Optional. Not available in China

2. Gantry

Feature	Specification
Aperture	720 mm
Rotation times	0.4*, 0.5, 0.75, 1.0, 1.5 seconds for full 360° scans
Focus-isocenter distance	570 mm
Focus-detector distance	1040 mm
Intercom system	Two-way connection between the gantry and console area
Gantry tilt	Scannable tilt range: -24° to 30° with 0.5° increments

* Optional

3. Patient table

Feature	Specification
Maximum scannable range	1,860 mm
Pitch	0.15 – 1.5
Z-position accuracy	+/- 1.0 mm
Longitudinal speed	1 mm/s – 300 mm/s
Lowest table height	530 mm
Maximum load capacity	452 lbs (205 kg)

4. Accessories

Accessories	
Knee pad	Standard
Neck cushion	Standard
Head rest cushion	Standard
Foot extension cushion	Standard
Patient straps and slideway unit	Standard
Head holder	Standard
Foot extension	Standard
Table pad	Standard
Flat head holder cushion	Standard
Arm support	Optional
Coronal head rest	Optional
Head coronal cushion	Optional
Flat head rest	Optional
Table side handle	Optional

5. Imaging chain

5.1 Generator

Feature	Specification
Effective power with iDose ⁴	72 kW, 94 kW*
Power rating	55 kW, 72 kW*
kVp setting	70, 80, 100, 120, 140
mA range (step size)	5 – 667 (1 mA step)

Note: Effective power is calculated by using generator power (55 kW, 72 kW*) and using iDose⁴ at the same time. This gives Incisive CT 64 Essentials effectively more power.

*Optional. Not available in China.

5.2 X-ray tube

Feature	Specification
Focal spot sizes, quoted to IEC 336/93 standard	Small: 0.5 x 1.0 Large: 1.0 x 1.0
Anode effective heat capacity	≥25 MHU _{eff} ; direct cooling
Anode heat capacity	8.0 MHU
Continuous anode input power (IEC 60613:2010)	≥5.5 kW
Anode diameter	200 mm
Anode rotation speed	105 Hz (6,300 rpm)
Target angle	7°
Maximum helical exposure time	120 s

5.3 NanoPanel Elite detector

Feature	Specification
Slices	64
Coverage	20 mm
Material	Solid-state GOS with 43,008 elements
Slip ring	Optical – 5 Gbps transfer rate
Data sampling rate	Up to 2,320 views/revolution/element
Collimations available	32 x 0.625 mm 2 x 0.625 mm 12 x 1.25 mm 16 x 0.625 mm 12 x 0.625 mm 4 x 0.625 mm
Slice thickness (helical mode)	0.67 mm – 5 mm
Slice thickness (axial mode)	0.625 mm – 10 mm
Scan angles	240°, 360°
Scan field of view	500 mm

6. Image quality

6.1 Spatial resolution

Spatial resolution	Specification
High mode	16.0 ± 10% lp/cm @ 0% MTF
Standard mode	13.0 ± 10% lp/cm @ 0% MTF

6.2 Low-contrast resolution

Feature	Specification
Low-contrast resolution	4 mm @ 0.3%; ≤ 15.5 mGy CTDI _{vol} (body)

Note: 20 cm Catphan phantom; 10 mm slice thickness

6.3 Other

Feature	Specification
Absorption range	-1,024 to +3,071 Hounsfield Units
Noise	≤ 0.18% at 120 kV, 230 mAs, 10 mm image thickness



7. Reconstruction

7.1 Reconstruction speed

Feature	Specification
Reconstruction speed	Up to 28 IPS

Optional

7.2 iDose⁴

iDose⁴ improves image quality* through artifact prevention and increased spatial resolution at low dose.

7.3 O-MAR

O-MAR reduces artifacts caused by large orthopedic implants.

7.4 Cone Beam Reconstruction Algorithm – COBRA

Philips patented Cone Beam Reconstruction Algorithm (COBRA) enables true three-dimensional data acquisition and reconstruction in both axial and helical spiral scanning.

7.5 ClearRay reconstruction

A revolutionary solution pre-computes and stores beam hardening and scatter corrections in a database later referenced to create a correction that is personalized to each individual patient. As a fully three-dimensional technique, contrast scale stability is preserved across different patient sizes, image uniformity is improved, and organ boundaries are better visualized.

7.6 Adaptive filtering

Adaptive filters reduce pattern noise (streaks) in non-homogenous bodies, improving overall image quality.

Optional

7.7 Adaptive multi-cycle reconstruction

Image data can be prospectively gated or retrospectively tagged. Adaptive multi-cycle reconstruction automatically delivers the best temporal resolution possible for the current scan.

7.8 Reconstruction field of view

50 to 500 mm

7.9 Image matrix

512 x 512 · 768 x 768 · 1,024 x 1,024

7.10 Off-line reconstruction

Off-line (batch) background image reconstruction of user-defined groups of raw data files with automatic image storage.



* Improved image quality is defined by improvements in spatial resolution and/or noise reduction as measured in phantom studies.

8. DoseWise

Philips DoseWise is a holistic approach to dose management that is active in every level of product design. It encompasses a set of techniques, programs and practices based on the ALARA (As Low As Reasonably Achievable) principle and supports outstanding image quality at low dose.

8.1 DoseRight Index

DoseRight Index (DRI) is a single number used to specify the image quality required for the diagnostic task at hand. DRI includes organ-specific DRI for the liver and the head and neck to provide appropriate dose and image quality within a single acquisition.

8.2 CT Dose Check

Supports an operator notification in each exam card that will be shown if an acquisition is planned that exceeds a specified $CTDI_{vol}$ or DLP. In addition, an alert is available such that, if an acquisition is planned and the total exam will exceed a specified $CTDI_{vol}$ or DLP, the operator will be required to enter his or her name and (if configured) a password to proceed, or the operator can adjust the scan parameters. Compliant with NEMA XR-25 and XR-29.

8.3 DICOM structured reporting IHE REM profile

DICOM radiation dose structured report that can be transferred to external systems such as HIS/RIS, PACS, or dose registries.

8.4 DoseRight automatic current selection

Personalizes dose for each patient by automatically suggesting tube current settings according to the estimated patient diameter in the scan region.

8.5 DoseRight angular dose modulation

Angular dose modulation varies the tube current during helical scans according to changes in patient shape (eccentricity) and tissue attenuation as the tube rotates. For each rotation, projections are processed to determine the maximum and minimum patient diameter. The tube current for the next rotation is then modulated between these limits.

8.6 3D-DOM

3D-DOM combines angular and longitudinal information to modulate dose in three dimensions.

8.7 Dedicated pediatric protocols

Age- and weight-based pediatric protocols produce diagnostic images at a minimum dose according to patient age for head scans, and patient weight for body scans.

8.8 Dose display and reports

Philips CT scanners include intuitive reporting and recording of estimated dose indices, dose reduction, and dose efficiency. Dose estimates are displayed on the operator's console for all scan protocols prior to and throughout the examination. Volume computed tomography dose index ($CTDI_{vol}$) and dose-length product (DLP) are automatically updated as the operator plans the scan. Also, a dose report may be included as a DICOM dose structured report and/or DICOM secondary capture with the reconstructed data set.

8.9 Dose performance data

$CTDI_{vol}$	Measurement
Head	14.07 mGy/100 mAs
Body	8.11 mGy/100 mAs

Measured on head and body $CTDI_{vol}$ phantoms (IEC 60601-2-44 ed. 3) at 120 kVp.

9. Clinical enhancements

Optional

9.1 Bolus tracking

An automated injection planning technique to monitor actual contrast enhancement and initiate scanning at a predetermined level.

9.2 Spiral Auto Start (SAS)

Spiral Auto Start allows the injector to communicate with the scanner. This allows the technologist to monitor the contrast injection and to start the scan (with a predetermined delay) while in the scan room.

9.3 Dental Application

Dental Planning generates panoramic views and cross-sectional cuts of mandible and maxilla from ultra-high-resolution axial slices.

9.4 Clinical applications – standard

- CT Reporting
- MPR
- MinIP
- Volume Rendering
- Filming
- MIP
- AIP
- Virtual Endoscopy

Optional

9.5 Brain Perfusion

Philips Brain Perfusion application allows the physician to differentiate areas of increased blood volume and decreased blood flow and presents this information that may help to distinguish between still-viable and non-viable infarcted tissue. Using serial CT scans obtained with intravenous injection of contrast, Brain Perfusion derives perfusion information from the time-density curves based on the uptake of injected contrast material and subsequent tissue enhancement. The application generates quantitative color maps of cerebral blood flow (CBF), cerebral blood volume (CBV), mean transit time (MTT) and time-to-peak (TTP).

9.6 Lung Nodule Assessment

The Lung Nodule Assessment provides nodule analysis and follow-up studies for doubling time to give reference details for benign and malignant definition of lung cancer, as well as early diagnosis and treatment.

9.7 CT Colonoscopy

The Virtual Colonoscopy can provide Automatic Colon Segmentation, virtual endoscopy, transparent view, cross sectional and unfold view. The Navigation Stage in viewer allows full freedom for clinicians to fly through the colon virtually and mark the polyps for measurements and clinical analysis.

9.8 Vessel Analysis

This set of tools for general vascular analysis allows the user to easily remove bone, and extract and segment the vessels to quickly perform typical measurements such as intraluminal diameter, cross sectional lumen area, and length of vessel's segments, and angle of the vessels. The package allows the user to display the data set using volume rendering, Average, or MIP with cross sections images that can be used to delineate aneurysm, presence of mural calcification and lining mural thrombus, branch vessel (celiac, mesenteric, renal) and the iliofemoral arterial runoff circulation.

9.9 Cardiology Solutions

Cardiac Plus Package

The Cardiac Plus package is a set of features designed to allow basic cardiovascular imaging of the heart. Includes Retrospective Tagging, Prospective Gating, Cardiac Calcium Scoring, Cardiac Coronary Artery Analysis and Cardiac Function Analysis.

9.10 Interventional Solutions

Interventional Essentials – Cart

Continuous CT (CCT) biopsy mode enables the clinician to perform scans from the gantry room using a foot pedal and view the images on a cart-mount, in-room monitor for guidance planning and monitoring.

9.11 Dual Energy

Includes a Dual Energy scan type that allows the acquisition and reconstruction of sequential dual-energy scans. The Spectral Analysis application may allow separation and analysis of materials such as calcium and uric acid when used with dual-energy scan data.

10. Networking and storage

10.1 Networking

Supports 10/100/1000 Mbps (10/100/1000 BaseT) networks. For optimal performance, Philips recommends a minimum 100 Mbps network (1 Gbps preferred) and for the CT network to be segmented from the rest of the hospital network.

10.2 DICOM

DICOM 3.0-compliant image format. Images can be auto-stored to selected archive media.

Includes the following DICOM functionality:

- Service-class user and profile (CT, secondary capture)
- DICOM Print
- DICOM Modality Worklist
- Query/Retrieve User
- Modality Performed Procedure Step User
- Removable Media
- Structured Reports

10.3 DICOM connectivity

Full implementation of the DICOM 3.0 communications protocol allows connectivity to DICOM 3.0-compliant scanners, workstations, and printers; supports IHE requirements for DICOM connectivity.

10.4 DICOM DVD/CD writer

Stores DICOM images and associated image viewing software on DVD/CD media. Images on these DVD/CDs can be viewed and manipulated on PCs meeting the minimum specifications. Suited for individual result storage and referring physician support.

10.5 Filming

Basic monochrome and color DICOM print capability are supported.

11. Site planning

11.1 Power requirements

- 200/208/240/380/400/415/440/460/480 VAC
- 50/60 Hz
- 115 kVA supply
- Three-phase distribution source
- Isolation transformer (optional)

Optional

11.2 Console Uninterrupted Power Supply (UPS)

Provides up to 30 minutes of backup power for host system.

11.3 Environmental requirements

Temperature

Gantry room	18 to 24°C (64 to 75°F)
Control room	10 to 30°C (50 to 86°F)
Operating maximum gradient	5 °C/hour
Storage temperature	-15° to 45°C (5 to 113°F)
Storage maximum gradient	10°C/hour
Transportation	-20°C to +50°C

Humidity


Gantry and control room	40% to 70%, non-condensing
Storage	20% to 80%, non-condensing
Transportation	20% to 80%, non-condensing

Air pressure

Operating	70 to 106 kPa
Storage	50 to 106 kPa
Transportation	50 to 106 kPa

11. Site planning continued

11.4 System requirements – minimum room layout



11.5 Dimensions and weights, per unit (without package)

Type	Length	Width	Height	Weight
① Gantry scanner	2,333 mm	973 mm	1,956 mm	1,995 kg
② Table	2,928 mm	700 mm	906 mm	360 kg
③ Scan control box	334 mm	117 mm	42 mm	2 kg
④ Console UPS (optional)	408 mm	430 mm	85 mm	16.8 kg
⑤ Isolation transformer (optional)	770 mm	592 mm	826 mm	370 kg



The Incisive CT is a computed tomography X-ray system intended to produce images of the head and body by computer reconstruction of X-ray transmission data taken at different angles and planes. These devices may include signal analysis and display equipment, patient and equipment supports, components and accessories. The CT Incisive 64 Essentials is indicated for head, whole body, cardiac and vascular X-ray computed tomography applications in patients of all ages.

These scanners are intended to be used for diagnostic imaging and for low dose CT lung cancer screening for the early detection of lung nodules that may represent cancer.* The screening must be performed within the established inclusion criteria of programs and protocols that have been approved and published by either a governmental body or professional medical society.

* Please refer to clinical literature, including the results of the National Lung Screening Trial (N Engl J Med 2011;365:395-409) and subsequent literature for further information.

The images and descriptions contained herein provide technical specifications and optional features which may not be included with the standard system configuration. Contact your local Philips representative for complete specific system details.

Some or all of the products, features, and accessories shown or described herein may not be available in your market. Please contact your local Philips representative for availability.

The Incisive CT 64 Essentials is a configuration of Incisive CT.



© 2020 Koninklijke Philips N.V. All rights are reserved. Philips reserves the right to make changes in specification and/or to discontinue any product at any time without notice or obligations and will not be liable for any consequences resulting from the use of this publication.

www.philips.com/Incisive-CT
4522 991 57111 * FEB 2020