

Clinical Bulletin #17

We are committed to supporting clinicians through the complete remediation process and will provide a range of resources to help you better inform, instruct, and support your patients. The clinical bulletin will continue to offer news and updates on the PE-PUR foam voluntary recall/FSN and will also be expanded to help address common topics and concerns on additional topics. Please find below some recent information that we wanted to bring to your attention.

Ventilation Updates

1. High flow therapy on the V60 and V60 Plus

Philips Respironics is complying with the FDA's Transition Plan for Medical Devices That Fall Within Enforcement Policies Issued During the Coronavirus Disease 2019 (COVID-19) Public Health Emergency guidance document, by removing the high flow therapy (HFT or HFT60) software option from all V60 and V60 Plus devices in the US.

Please note: use of High Flow Therapy (HFT) on the V60 or V60 Plus ventilators after November 7, 2023 will be considered off-label use.

For more information, please visit www.Philips.com/hft-removal.

2. OmniLab Advanced + Remediation Update

This month, we will be reaching out to customers who own OmniLab Advanced + sleep lab devices. Customers who own registered OmniLab Advanced + devices will be contacted directly with a ship-to form that will list all registered affected OmniLab Advanced + serial numbers and allow the customer to choose between 2 options: receive a credit within 6 weeks for each affected device or continue to wait for additional remediation options to become available. Rental devices will be addressed at a later date.

Please reach out to your Philips Respironics Account Manager for any additional questions.

Useful Links

PE-PUR Foam voluntary recall/FSN - Information for Clinicians

Be sure to visit [our clinicians information page](#) regularly for the most current information for you and your patients.

Masks and Magnets

Have additional questions about Philips Respironics masks and magnets? Learn more [here](#).

Ventilation

For the latest ventilation news and updates, please visit [this page](#).



For more information, visit Philips.com/src-update

