

Clinical Bulletin #18

We are committed to supporting clinicians through the complete remediation process and will provide a range of resources to help you better inform, instruct, and support your patients. The clinical bulletin will continue to offer news and updates on the PE-PUR foam voluntary recall/FSN and will also be expanded to help address common topics and concerns on additional topics.

Please find below some recent information that we wanted to bring to your attention.

Ventilation Updates

1. V30

As of October 31, 2023 Philips has discontinued the sale of the Philips Respironics V30 Auto BiPAP device due to product lifecycle planning.

Part #	Product Description
1135427	BiPAP V30 Auto mobility bundle (includes V30 Auto, roll stand, detachable battery pack and detachable battery pack enclosure)
1111178	BiPAP V30 Auto (includes ventilator, power cord, and clinical manual)
1143671	BiPAP V30 Auto Core

We will continue to honor valid service contracts and warranty claims. As such, we will make service parts and bench repair services for the V30 Auto device available for purchase for five years (through December 2028), as long as components remain available.

2. High flow therapy on the V60 and V60 Plus

As a reminder, High Flow Therapy (HFT) on the V60 or V60 Plus ventilators after November 7, 2023 is not to be used.

Philips Respironics is complying with the FDA's Transition Plan for Medical Devices That Fall Within Enforcement Policies Issued During the Coronavirus Disease 2019 (COVID-19) Public Health Emergency guidance document, by removing the high flow therapy (HFT or HFT60) software option from all V60 and V60 Plus devices in the US.

For more information, please visit www.Philips.com/hft-removal.

Useful Links

PE-PUR Foam voluntary recall/FSN - Information for Clinicians

Be sure to visit our [clinicians information page](#) regularly for the most current information for you and your patients

Masks and Magnets

Have additional questions about Philips Respironics masks and magnets? Learn more [here](#).

Ventilation

For the latest ventilation news and updates, please visit [this page](#).



For more information, visit Philips.com/src-update