

Lumify iOS Device Compatibility List

Supported Devices*

Important notice: For iOS USB-C devices users, please turn off the auto-upgrade on your iOS devices and do not update to iOS18 or later due to an OS issue Philips team is investigating.

Items in RED are new to this document revision Items in Blue are compatible, but no longer available								
Device Name	Year Released	Screen Size (in.)	Connectivity Options	Internal Storage (in GB)	Philips Case Available	LPM Required	Connector Type	Supported OS
iPad Air 11-inch M2	2024	11	Wi-Fi / LTE	128	No	Yes	Type C	17.5.1
iPad Air 13-inch M2	2024	13	Wi-Fi / LTE	128	No	Yes	Type C	17.5.1
iPad Pro 12.9-inch 6th Gen	2022	12.9	Wi-Fi / LTE	128	No	Yes	Type C	17.3
iPad Pro 11- inch (M4)	2024	11	Wi-Fi / LTE	256	No	Yes	Type C	17.5.1
iPad Air 5th Gen M1	2022	10.9	Wi-Fi / LTE	64	No	Yes	Type C	17.2
iPad Pro 13- inch M4	2024	13	Wi-Fi / LTE	512	No	Yes	Type C	17.5.1
iPhone 15	2023	6.12	Wi-Fi / LTE	128	No	Yes	Type C	17.0.3
iPhone 15 Plus	2023	6.69	Wi-Fi / LTE	128	No	Yes	Type C	17.0.3
iPhone 15 Pro	2023	6.12	Wi-Fi / LTE	128	No	Yes	Type C	17.2.1
iPhone 14	2022	6.06	Wi-Fi / LTE	128	No	Yes	Lightning	16.6
iPhone 14 Pro	2022	6.1	Wi-Fi / LTE	128	No	Yes	Lightning	16.3.1
iPad 10th Gen	2022	10.9	Wi-Fi / LTE	64/256	No	Yes	Type C	16.3.1
iPad Pro 6th Gen	2022	12.9	Wi-Fi / LTE	128/256/512 /1TB/2TB	No	Yes	Type C	16.5
iPad Air 5th Gen	2022	10.9	Wi-Fi / LTE	64/256	No	Yes	Type C	16.5



Lumify iOS Device Compatibility List

iPad Mini 6th Gen	2021	8.3	Wi-Fi / LTE	64/256	No	Yes	Type C	16.6
iPhone 13 mini	2021	5.4	Wi-Fi / LTE	128/256/512	No	Yes	Lightning	Out of the box
iPhone 13	2021	6.1	Wi-Fi / LTE	128/256/512	No	Yes	Lightning	Out of the box
iPad (9th gen)	2021	10.2	Wi-Fi / LTE	64/256	Yes	Yes	Lightning	Out of the box
iPhone 12	2020	6.1	Wi-Fi / LTE	64/128/256	No	Yes	Lightning	Out of the box

***Disclaimer:**

Devices not included in the preceding table have not undergone compatibility testing with Lumify. Philips cannot guarantee Lumify system performance with non-compatible devices as compatibility testing with all devices on the market is not feasible. The devices have been evaluated using the operating system (OS) versions they were originally equipped with.

Subsequent OS releases, known as maintenance releases, that are higher than the version installed at the time of device launch has not been explicitly verified. Conducting explicit tests for newer major OS releases on older devices, which were initially equipped with a previous OS version, is not feasible. Lumify system functionality and performance may be impacted if significant operating system changes have been applied.

Please note that Reacts tele-ultrasound solution is not available on Lumify when using the above iOS devices. All iOS devices currently compatible require use of Lumify Power Module and utilize USB-C transducer cables and will all continuously scan for 2 to 2.5 hrs.

Devices on the supported list satisfy all the following criteria: 1. Passed software testing on the latest Lumify software, 2. Passed Philips hardware testing, 3. Is readily available from the device manufacturer.

Philips cannot test every device available on the market, so this list represents a range of devices to satisfy a variety of user needs with respect to form factor, price, battery life and connectivity options.

